

INSTRUCTIONS FOR USE

THREE WAY STOP COCK WITH EXTENSION TUBE

Materials Used

- Polypropylene, Poly Vinyl Chloride (PVC), Polycarbonate, Acrylonitrile Butadiene Styrene (ABS), HDPE.

Indications

- Infusion of multiple I.V. fluids, parenteral nutrition and administration of other drugs and medicines.
- Maintaining hydration in patients who are unable to take sufficient volume of oral fluids.

Contraindications

- Do not use Non-Lipid 3-way stop cock with Lipid fluids, otherwise crack problem in the component of the product may occur.
- Not to be used in patients with known hypersensitivity to any of the materials used.
- Blood transfusion.

Directions for Use

- Remove the protective cap of male luer connector.
- Connect the female luer connector of extension tube to the male fitting of infusion set.
- Open the regulating clamp of I.V. set. Expel the air from the tube and close the regulating clamp and connect the extension tube with I.V. device
- Regulate the flow by adjusting regulating clamp to achieve desired flow rate.
- Open the white cap of three way stop cock and connect with the primed IV set.

Cautions

- Do not use if package is opened or damaged.
- Read carefully instructions for use before using the product.
- The product should be used only by qualified healthcare professionals.
- Do not use stopcock with extension line at more than 2 bar liquid pressure.
- Some organic solvents, alcoholic disinfectants, infusion solutions and high pH-value substances might cause cracking of the device. It is recommended to follow the IFU of such medicine / disinfectant before using with the device.
- Do not re-sterile may increase the EO Residual in the product or physical property of the products may be change.
- Product is single use only. Re-use of the product can cause change in physical characteristics like handle loose, crack problem, Blockage, Infection & other blood related transfusion diseases like HIV, AIDS & Hepatitis etc.
- Reuse and cleaning of product may alter their structural and mechanical properties. It may lead to infection or other illness/injury.
- The product should be replaced and disposed of as per facility approved protocol or CDC guidelines.
- Store in between 5°C to 35°C, avoid excessive heat, protect from direct sunlight and moisture.
- Poly Medicure Limited will not be responsible for any direct incidental or consequential damages resulting from reuse of product.



POLY MEDICURE LTD.

Plot No.104-105, Sector-59, HSIIDC, Industrial Area,
Ballabgarh Faridabad, Haryana-121004, INDIA.
Web : www.polymedicure.com,
Contact customer care executive: +91-129-4287053,
Email: customercare@polymedicure.com,
Address: Same as above

Obelis S.a
Boulevard Général Wahis 53
1030 Brussels, BELGIUM
Tel : +(32) 2. 732.59.54
E-mail : mail@obelis.net



Cautions



See Instructions for use



Product reference/Art. No.



For single use only



Do not Resterilize



Do not use if packaging or product has been damaged or contaminated



Batch Number



Date of Manufacturing



Use by / Expiry date



Sterilised by Ethylene Oxide



Single sterile barrier system



European Authorised Representative



Manufactured by



Non pyrogenic



Storage Condition



Keep Away from Sunlight



Keep Dry



Medical Devices