

# INSTRUCTIONS FOR USE

## I.V. Infusion Set (Vented) **PHOTOFUSION SET**

### MATERIALS USED :

- Polypropylene, Acrylonitrile Butadiene Styrene (ABS), Poly Vinyl Chloride (PVC)

### INDICATIONS :

- Infusion of photo sensitive drugs .

### CONTRAINdications :

- Administration of highly viscous fluids.
- Blood Transfusion.
- Not to be used in patients with known hypersensitivity to any of the materials used.

### DIRECTIONS FOR USE :

- Close the Regulating Clamp.
- Remove the protective cap from the spike and insert the spike into the I.V. fluid container to its full length.
- Squeeze the drip chamber and adjust the fluid level about 2/3rd filled.
- Open the air vent cap of piercing spike.
- Open the regulating clamp, priming takes place automatically up to integrated filter (maintains a close system thus prevents contamination). Close the regulating clamp.
- Remove the priming filter cap.
- Connect the IV set with I.V. device and regulate the flow by adjusting regulating clamp to achieve desired flow rate.

### CAUTIONS :

- Do not use if protective cap is loose or missing.
- Do not use if package is opened or damaged.
- For gravity feed only.
- Close the airvent during periods of interrupted infusion therapy.
- The product should be used only by qualified healthcare professionals.
- Do not re-sterilize. Discard the set after single use.
- Reuse and cleaning of product may alter their structural and mechanical properties. It may lead to infection or other illness/injury.
- The product should be replaced and disposed of as per facility approved protocol or CDC guidelines.
- Store in between 5°C to 35°C, avoid excessive heat, protect from direct sunlight and moisture.
- Poly Medicure Limited will not be responsible for any direct incidental or consequential damages resulting from reuse of product.

 **POLYMED**

 **POLY MEDICURE LTD.**

Plot No. 17, Sector-3, Integrated Industrial Estate,  
SIDCUL, Hardiwar, Uttrakhand 249403, INDIA,

Web : [www.polymedicure.com](http://www.polymedicure.com)

Contact customer care executive: +91-129-4287053,

Email: [customercare@polymedicure.com](mailto:customercare@polymedicure.com)

Address: Same as above

Not made with DEHP plasticizer

 **Obelis s.a.,**  
Boulevard Général Wahis 53,  
1030, Brussels, BELGIUM.  
Tel : +(32) 2. 732.59.54  
E-mail : [mail@obelis.net](mailto:mail@obelis.net)

 **0123**



Cautions



See Instructions for use



Product reference/Art. No.



For single use only



Do not Resterilize



Do not use if packaging or product has been damaged or contaminated



Batch Number



Date of Manufacturing



Use by / Expiry date



Sterilised by Ethylene Oxide



Single Sterile Barrier System



European Authorised Representative



Manufactured by



Non-Pyrogenic



Storage Condition



Keep Away from Sunlight



Keep Dry



Medical Devices